

645—63.12(157) Disinfecting nonelectrical instruments and equipment.

63.12(1) Before use upon a client, all nonelectrical instruments shall be disinfected by an EPA-registered, hospital-grade disinfectant with demonstrated bactericidal, fungicidal, and virucidal activity and used according to the manufacturer's instructions.

- a.* All disinfected instruments shall be stored in a clean, covered place.
- b.* All instruments that have been used on a client or soiled in any manner shall be placed in a proper receptacle.
- c.* Disinfectant solutions shall be changed at least once per week and whenever visibly cloudy or dirty.

63.12(2) If the nonelectrical instruments and equipment specified in this rule are sterilized in accordance with the requirements outlined in rule 645—63.15(157), the requirements of this rule shall be fulfilled.